NASA CR-54439

FIFTH QUARTERLY REPORT

PRESSURE MEASURING SYSTEMS FOR

CLOSED CYCLE LIQUID METAL FACILITIES

Anthony J. Cassano

PREPARED FOR

NATIONAL AERONAUTICS AND SPACE ADMINISTRATION

CONTRACT NAS 3-4170

June 28, 1965

Consolidated Controls Corporation

Bethel, Connecticut

A SUBSIDIARY OF **CONDEC** CORPORATION

NOTICE

This report was prepared as an account of Government sponsored work. Neither the United States, nor the National Aeronautics and Space Administraion (NASA), nor any person acting on behalf of NASA:

A.) Makes any warranty or representation, expressed or implied, with respect to the accuracy, completeness, or usefulness of the information contained in this report, or that the use of any information, apparatus, method, or process disclosed in this report may not infringe privately owned rights; or

B.) Assumes any liabilities with respect to the use of, or for damages resulting from the use of any information, apparatus, method or process disclosed in this report.

As used above, "person acting on behalf of NASA" includes any employee or contractor of NASA, or employee of such contractor, to the extent that such employee or contractor of NASA, or employer of such contractor prepares, disseminates, or provides access to, any information pursuant to his employment or contract with NASA, or his employment with such contractor.

Requests for copies of this report should be referred to:

National Aeronautics and Space Administration
Office of Scientific and Technical Information
Washington, D. C. 20546
Attention: AFSS-A

FIFTH QUARTERLY REPORT
March 1, 1965--May 31, 1965

PRESSURE MEASURING SYSTEMS FOR CLOSED CYCLE LIQUID METAL FACILITIES

PREPARED FOR

NATIONAL AERONAUTICS AND SPACE ADMINISTRATION

CONTRACT NAS 3-4170

JUNE 28, 1965

Technical Management
NASA - Lewis Research Center
Nuclear Power Technology Branch
R. N. Weltmann

Prepared by:

Anthony J. Cassano
Research Engineer

Approved:

R. E. Engdahl
Project Manager

Consolidated Controls Corporation
Bethel, Connecticut

FOREWORD

The major contributors to this development

program are Mr. R. Engdahl, Project Manager,

Mr. Anthony Cassano, Mr. David Mends and Mr.

Philip Tubman.

ABSTRACT

28807

Continuing development of a pressure transducer system
for liquid metal applications is described. During the
report period, emphasis was placed upon fabrication of
an experimental transducer incorporating a thermionic
diode sensor. Problems encountered in fabricating the
transducer are outlined along with the anticipated
solutions. A potassium compatibility test facility has
been constructed and checked out. A complete set of
single convolution pressure capsules and electrical
terminals has been charged with liquid potassium and is
undergoing the initial stages of the compatibility test
procedure. Preliminary analyses are presented for
different configurations of differential pressure
transducers.

i

TABLE OF CONTENTS

LIST OF ILLUSTRATIONS

1.0 Introduction

The objective of this program is to develop pressure transducers which can be used in advanced closed cycle power systems using liquid metals such as mercury, sodium, potassium and other alkali metals as working and heat transfer media at elevated temperatures. Accurate pressure measurements in the high temperature liquid, vapor, and two phase streams are required for research, design and control purposes. In addition, space flight requires light weight systems capable of enduring long periods of unattended operation.

Liquid metal pressure measurements at elevated temperatures pose many design problems demanding the best from available materials. To establish a firm design base for the transducer equipment, four diaphragm materials and two transducer systems have been chosen for evaluation. The selected transducer system using a thermionic diode sensor will be developed for use as either ground or flight hardware for measuring absolute and differential pressures. The absolute pressure instrument will be developed for a full scale range of 80 psia and the differential instrument for ±5 psid.

2.0 Summary

During the report period, emphasis was placed on the fabrication of an experimental transducer incorporating a thermionic diode sensor. The main problem areas involve the encapsulated heater assembly and the technology required to obtain reliable metal-ceramic seals. Initial difficulties encountered with the encapsulated heater assembly have been alleviated by the introduction of a flat alumina coil form on which the heater is wound, and the use of alumina insulators in the casting cavity to avoid shorting of the heater to the emitter housing. Present plans also call for the introduction of a rhenium heater coupled with rhenium heater power leads to improve high-temperature reliability.

The pressure-temperature life test on a sample metal-ceramic seal planned for use in the transducer design was concluded after 2400 hours at 1800°F, the last 1000 hours of which were at 1000 psia pressurization. The sample employed Lucalox/Nb-1Zr seals fabricated by use of a tungsten/yttria metallizing

2.

and a nickel/niobium/titanium braze material. This technique is being used in the fabrication of the experimental transducer discussed above. Efforts are being made to obtain a meaningful metallographic analysis of the test sample.

Construction and initial check out of the potassium compatibility test chamber was completed. A complete set of single convolution pressure capsules and electrical terminals has been charged with liquid potassium, installed in the compatibility test chamber and is undergoing the initial stages of the compatibility test procedure. It is planned to expose the test capsules to liquid potassium at 1800°F (corresponding to a potassium vapor pressure of about 80 psia) for a series of five 100 hour test cycles. This will result in a total of 500 test hours, at the conclusion of which the test capsules will be examined for possible potassium attack.

Preliminary analyses were performed on possible configurations for the differential pressure

transducers. Results indicate the feasibility of obtaining 0.001 inch motion of the active collector of a thermionic diode sensor. Work has started on the design and assembly of a mock-up configuration to verify the results of the analysis.

4.

3.0 Thermionic Diode Pressure Transducer

A preliminary design for a complete pressure transducer was presented in the Fourth Quarterly Report (Reference 1). The design, incorporating a double convolution pressure capsule and a thermionic diode sensor, is shown in Figure 1. Work on the first experimental device has concentrated on the fabrication of the encapsulated heater assembly, the preparation of its ceramic base and the development of the metal-ceramic technology necessary to obtain the required seal integrity. For the first device, the heater assembly will be mounted on its ceramic base, the ceramic base will be brazed into a Nb-1Zr support plate and the entire assembly will be installed in a micrometer head fixture (Figure 9, Reference 2; Figure 15, Reference 3) which allows controlled movement of a simulated active collector

Initial efforts to fabricate the encapsulated heater assembly utilized a bare tungsten wire of 0.003 inch diameter spirally wound on a 0.004 inch diameter

mandrel. The resultant heater had an outer diameter of 0.010 inch and a length of 0.875 inch. The heater was threaded through alumina insulating tubes and cast into the molybdenum emitter housing with an alumina/ beryllia mixture. The heater power leads were 0.010 inch diameter molybdenum wire connected to the heater by plasma-spraying the joint with tungsten and embedding the joint in the casting. Three difficulties were noticed with this technique.

1.   The threading of the heater through the alumina tubes to form a module suitable for insertion into the emitter housing cavity proved to be a delicate and time-consuming operation.

2.   Four or five separate casting operations were required to completely fill the voids left after the heater had been inserted into the cavity. In addition to the space remaining around the heater, the casting material must also fill the void space inside the alumina

6.

insulating tubes. After each cast (3550°F), a certain amount of motion was noticed in the heater wire, apparently due to softening of the alumina tubes at the casting temperature. The wire motion resulted in shorting of the heater to the emitter housing. In two instances, large portions of the heater were shorted in this manner, resulting in the loss of the entire assembly.

3. Mounting the encapsulated heater assembly on its ceramic base requires that sharp bends be made in the molybdenum heater power leads so that the wires may be inserted in the proper holes in the ceramic base. However, after the casting operations, the molybdenum wires become brittle and thus subject to breakage during the bending operation. In this event, a 0.010 inch diameter tantalum extension lead was welded to the molybdenum stub and brought through the holes in the ceramic base.

7.

The technique now under development for building the
heater is designed to alleviate the difficulties
presented above. The alumina insulating tubes have
been replaced by a flat grooved alumina coil form on
which the heater wire is wound. The void spaces
inside the tubes are thereby eliminated. To further
eliminate voids in the cavity and to prevent the
heater shorting to the emitter housing during the
casting, pieces of alumina are placed between the
heater and the walls on all four sides and at the
bottom of the cavity. In addition, by minimizing
the voids, a reduction in the total number of casting
operations needed to fill the cavity was possible.

The above described heater assembly is shown in
Figures 2 and 3. The basic heater assembly, consist-
ing of the alumina coil form, the bare tungsten
heater and the molybdenum power leads, is shown in
Figure 2. This basic heater assembly is then
inserted in the emitter housing cavity along with
the alumina insulating plates mentioned above.
Figure 3 shows this assembly prior to the casting

8.

operation. The casting temperature (3550°F) prohibits the installation of the actual emitters until the casting is completed. The cavities for the two emitters are shown in Figure 3.

The first test device was constructed using the heater assembly of Figure 3 in conjunction with the ceramic base shown in Figure 4. The complete assembly, with the Nb-1Zr support plate, is shown in Figure 5. Three metal-ceramic joints were used in the complete assembly. These include:

1. Mounting of the encapsulated heater assembly (molybdenum) to the Lucalox ceramic base,
2. Mounting of the Nb-1Zr support plate to the Lucalox ceramic base and
3. Mounting of the reference collector surface to the Lucalox ceramic base.

All the above joints must, of course, maintain mechanical stability. In addition, joining operations (1) and (2) must take place at a low enough temperature

9.

(below 2300°F) to prevent damage to the thermionic
emitters. The reference collector may be mounted
at a higher temperature since this operation takes
place prior to mounting of the encapsulated heater
assembly (see Figure 4). For the first test assembly,
tungsten/yttria was used (1) as the metallizing
material for the posts on which the encapsulated
heater assembly was brazed, (2) as the metallizing
material for the braze between the ceramic base and
its Nb-1Zr support plate and (3) to establish the
reference collector surface. A nickel/niobium/
titanium braze material was used to make both the
heater assembly/ceramic and the ceramic/support plate
joints.

The assembly of Figure 5 was mounted on the vacuum
flange equipped with the micrometer head, installed
in a chamber of the Vacuum Test Facility and baked-out
to 1800°F with a chamber pressure of 3 X 10^{-8} torr.
At this point, the encapsulated heater was energized

and emitter activation was begun. After about 16 hours
of operation, during which time the emission currents
from both emitters were close to the values predicted
by space-charge theory, the heater developed an open
circuit. Preliminary examination indicated that the
open circuit occurred in the first heater turn around
the alumina coil form, which includes one of the plasma-
sprayed joints. Efforts are being made to disassemble
the heater to determine exactly where the failure
occurred.

Present plans call for the use of rhenium wire instead
of tungsten wire to improve the high-temperature
characteristics of the heater and its external leads.
The use of rhenium should bring about the following
improvements.

1. A rhenium/rhenium weld joint will replace the
    tungsten/molybdenum plasma-sprayed joint now in
    use. A stronger, more stable connection should
    result.

2. The improved ductility of rhenium will minimize
    the breakage problem caused by bending the
    external heater leads after casting.

11.

At present, 0.010 inch diameter rhenium wire has been obtained and will be used for the external leads.

4.0 Metal-Ceramic Seal Tests

The thermionic diode pressure transducer requires the use
of metal-ceramic seals capable of maintaining vacuum
integrity under high-temperature conditions. In the
event of pressure capsule failure, the metal-ceramic
seals must also be able to withstand liquid metal
attack. At present, two approaches are being evaluated,
both of which show promise of being satisfactory seal
techniques.

1. A group of electrical terminal configurations
    (Figure 13, Reference 2) was obtained for
    evaluation. The terminals employed Nb-1Zr/
    Lucalox seals. One of the terminals was
    subjected to three consecutive test cycles with
    the following parameters.

    400 hours, 1800°F, 300 psia external argon pressure

    400 hours, 1800°F, 600 psia external argon pressure

    400 hours, 1800°F, 1000 psia external argon pressure

    No electrical or mechanical failure was observed
    during the course of the test (Reference 3).

Four of the group of terminals were prepared
for, and at present are undergoing, potassium
compatibility testing (see Section 5.0).

2. A life test was performed to evaluate the metal-
    ceramic joining technique presently being used
    to fabricate the initial thermionic diode
    pressure transducer models. The test sample
    (Figure 87, Reference 1), as with the electrical
    terminals mentioned above, contained Nb-1Zr/
    Lucalox seals. The Lucalox was metallized with
    a tungsten-yttria compound and a nickel/niobium/
    titanium braze material was used. The test
    schedule was as follows.

1000 hours, 1800°F, 500 psia internal argon pressure
 100 hours, 1800°F, 600 psia internal argon pressure
 100 hours, 1800°F, 700 psia internal argon pressure
 100 hours, 1800°F, 800 psia internal argon pressure
 100 hours, 1800°F, 900 psia internal argon pressure
1000 hours, 1800°F, 1000 psia internal argon pressure

14.

This test represents a more stringent evaluation
than that performed on the terminals described
above. Internal pressurization was used,
resulting in tensile stress on the seals.

The external pressurization applied to the
terminal configuration produced compressive
stress components which tended to maintain the
seal integrity.
The test was terminated by a failure in the weld
between the argon pressurization tubing (Nb-1Zr)
and a Nb-1Zr disk, not involving the metal-ceramic
joints under test. No electrical or mechanical
failure was observed during the test. Efforts are
being made to obtain a meaningful metallographic
analysis of the test unit. Present plans call for
preparation of test units for inclusion in the
potassium compatibility test program.

5.0 Compatibility Test Program

The compatibility test chamber, with its associated heaters, shields and cooling water coils, was completed and installed on its vacuum system (100 liters per second ion type pump). Initial testing of the heaters and bake-out of the chamber internal surfaces were completed. The chamber temperature was monitored by two thermocouples. At 1800°F, there was a difference of less than 5°F between the readings of the thermocouples. At 1800°F, a chamber pressure of 5×10^{-7} torr was obtained; at room temperature, the chamber pressure dropped to 10^{-8} torr. Figure 6 shows the internal components of the test chamber. Figure 7 shows the test chamber in operation with its vacuum system.

Four single convolution test capsules of C-129Y, FS-85, T-222 and W-25 Re alloy and one test capsule containing four representative electrical terminals have been charged with liquid potassium and installed in the compatibility test chamber. The test capsules

16.

were suspended in the chamber from the capsule support plate shown in Figure 6. The test capsule design configurations are described in the Third Quarterly Report (Figures 54, 55 and 56, Reference 3). Figure 8 is a photograph of the set of completed test capsules.

Chemical analyses have been obtained of the potassium used to fill the five (5) test capsules. Samples of the potassium were cast in a stainless steel tube under a vacuum of 5×10^{-5} torr in an electron beam welding chamber, subsequent to filling the last capsule. The potassium samples were analyzed for oxygen by the mercury amalgamation method (helium cover gas) and for metallic impurities by spectrographic techniques. The results are as follows:

Oxygen as K_2O: 3.4/7.6 ppm

Metallic Impurities in KCl:

    Ag, Al, Cb, Co, Cr, Cu, Fe, Mg,
    Mn, Mo, Ni, Pb, Sn, Ti, Zr: less than 1 ppm

    Ca: 1 ppm

    V: less than 5 ppm

    Na, Si: 5 ppm

A test program has been developed to evaluate the
potassium compatibility of the four candidate alloys
and the metal-ceramic seals. The program includes
the preparation of the test capsules, the actual
test procedures, and post-mortem testing. The
program is defined as follows.

1. Preparation of Test Capsules

    a. Obtain a chemical analysis of the capsule
        materials.

    b. Obtain a chemical analysis of potassium
        batch.

    c. Fill the test capsules with potassium and
        identify the capsules.

    d. Weigh the completed test capsules.

2. Test Procedure

    a. Install the test capsules into the
        compatibility test chamber.

    b. Raise test chamber temperature to 1800°F.
        At no time shall the chamber pressure
        exceed 10^{-5} torr. For chamber temperatures
        above 700°F, the chamber pressure shall not
        exceed 5×10^{-7} torr.

18.

c. Maintain test chamber temperature at 1800°F
    for a period of 100 hours. This will
    establish a potassium vapor pressure inside
    the test capsules of about 80 psia. After
    100 hours, bring the test chamber to room
    temperature.

d. Perform 2.c above five times to accumulate
    a total of 500 hours operation at 1800°F.

e. Remove the test capsules from the test
    chamber.

3. Post-Mortem Tests

a. Weigh the test capsules.

b. Visually inspect the test capsules for
    possible failures.

c. Extract the potassium from the test capsules
    and perform a chemical analysis on the
    potassium.

d. Test the capsules for leakage on a mass
    spectrometer.

e.   Cut up, section and perform metallographic examination of diaphragms, weld joints, braze joints, ceramics, etc.

f.   Perform chemical or microprobe analysis of any corroded layers indicated in 3.e above.

At present, the five (5) test capsules are starting the second 100 hour test cycle outlined in 2.c and 2.d above. At the conclusion of the first 100 hour test cycle, the chamber pressure was 5×10^{-8} torr. Upon returning the chamber to room temperature, the pressure dropped to 6×10^{-10} torr. It is expected that the time needed to raise the chamber temperature to 1800°F, subject to the pressure requirements oi 2.b above, will decrease as operating time at 1800°F is accumulated. The chamber pressure at 1800°F should also decrease for each successive 100 hour test cycle as the chamber becomes more thoroughly out-gassed.

6.0   Differential Pressure Transducer

Preliminary analyses have been performed on three mechanical configurations for the differential pressure transducer (±5 psid). Figure 9 shows the three configurations schematically. Each configuration utilizes the deflection of a composite beam assembly consisting of (1) an outer tubular beam member and (2) an internal beam member whose termination acts as the active collector surface for the thermionic diode sensor. The outer beam member will be exposed to the liquid potassium environment. A wall thickness of 0.025 inch has been assumed to be sufficient for liquid metal containment. The internal beam member will be sealed from the liquid potassium and connected to the vacuum environment of the thermionic diode sensor.

The envisioned differential pressure transducer is of the same basic design as shown in Figure 1. For the differential pressure application, the thermionic diode sensor will be removed from the interior of the

pressure capsule, isolated from the pressurized

volumes, and probably embedded in the capsule hous-

ing. The pressures P_1 and P_2, whose differential

$(P_1 - P_2)$ is to be monitored, are applied to the

external and internal surfaces, respectively, of

the pressure capsule. The main capsule housing

inner diameter is assumed to be about 1.5 inches

and the active collector surface extends 0.375 inch

beyond the fixed support of the outer beam member.

For Case I (cantilever, end load), the displacement of

the pressure capsule is transmitted directly to the canti-

levered end of the outer beam member, establishing an

end slope θ_1 relative to the horizontal position.

The deflection of the active collector surface from

equilibrium therefore becomes $(0.75 + 0.375)\,\theta_1$

inch less the deflection of the outer beam end. For

Case II (end couple, one end fixed, one end supported),

the pressure capsule displacement is transmitted

by a yoke arrangement as a couple acting on a

22.

supported pivot point at one end of the outer beam member. If the end slope at the supported pivot point is θ_2, the active collector deflection from equilibrium becomes $(1.5 + 0.375)\theta_2$ inch. Case III (cantilever, combination end load and end couple) involves the removal of the supported pivot point of Case II, allowing free movement of the outer beam member subject to the superposition of two actions; i.e., the pressure capsule deflection transmitted by the yoke arrangement as (1) an end load and (2) an end couple.

The applicable relations are as follows (Reference 4).
Case I (cantilever, end load)

$$\theta_1 = -\frac{W t^2}{2EI} \tag{1}$$

$$y_1 = \frac{W t^3}{3EI} \tag{2}$$

Case II (end couple, one end fixed, one end supported)

$$\theta_2 = \frac{MT}{4EI} = \frac{Wt(2t)}{4EI} = \frac{Wt^2}{2EI} \tag{3}$$

23.

Case III (cantilever, combination end load and end couple)

Due to end load:

$$\theta_{3L} = -\frac{WT^2}{2EI} = -\frac{2Wt^2}{EI} \tag{4}$$

$$y_{3L} = \frac{WT^3}{3EI} = \frac{8Wt^3}{3EI} \tag{5}$$

Due to end couple:

$$\theta_{3C} = \frac{MT}{EI} = \frac{Wt(2t)}{EI} = \frac{2Wt^2}{EI} \tag{6}$$

$$y_{3C} = -\frac{MT^2}{2EI} = -\frac{Wt(4t^2)}{2EI} = -\frac{2Wt^3}{EI} \tag{7}$$

Combining end load and end couple effects

$$\theta_3 = \theta_{3L} + \theta_{3C} = 0 \tag{8}$$

$$y_3 = y_{3L} + y_{3C} = \frac{2Wt^3}{3EI} \tag{9}$$

In these equations, the sign convention employed is that end slopes and deflections are positive when in the upward direction (direction of force W in Figure 9) and

θ_1 = tubular outer beam end slope (Case I) (radians)

θ_2 = tubular outer beam end slope (Case II) (radians)

24.

θ_3 = tubular outer beam end slope(Case III)(radians)

y_1 = tubular outer beam end deflection (Case I)(inch)

y_3 = tubular outer beam end deflection(Case III)(inch)

W = force exerted by pressure capsule on beam
(assumed to be 1 pound)

t = tubular outer beam length (Case I) (0.75 inch)

T = tubular outer beam length(Cases II & III)(1.5 inch)

M = moment produced by pressure capsule (Cases II & III)
(Wt pound-inches)

E = modulus of tubular outer beam material
(assumed to be 30×10^6 psi)

I = moment of inertia of tubular outer beam section
= $0.049 (D^4 - d^4)$ inch^4

D = outer diameter of tubular outer beam (inch)

d = inner diameter of tubular outer beam (inch)

The results of the preliminary analysis may be presented

as follows.

Tubular outer beam O.D. (inch)	0.25	0.20	0.15
I.D. (inch)	0.20	0.15	0.10

Active collector deflection
(10^{-3} inch)

	Case I	-0.052	-0.109	-0.294
	Case II	0.155	0.328	0.884
	Case III	0.083	0.175	0.472

25.

The data show that Case II holds promise of providing
an active collector deflection of 0.001 inch for
reasonably sized tubular beams. The value of 0.001 inch
full scale deflection has been found to result in a
feasible output for the thermionic diode sensor
(Figure 83, Reference 4). The difficulties inherent
in Case II, which requires that the supported pivot
point be inside the main capsule housing, make it
mandatory to concentrate further analysis on Cases I
and III, for which the fabrication requirements
would be far less stringent. Case III has the further
advantage of providing a purely vertical active
collector travel, since $\theta_3 = 0$. Work is in
progress to fabricate a mock-up of the Case III
configuration to verify the results of the analysis.

26.

PRESSURE
PORT

TWO
CONVOLUTION
CAPSULE

DUAL
EMITTERS

ACTIVE
COLLECTOR

EVACUATION
& PINCH-OFF
TUBE

ADJUSTING
RING

ACTUAL
SIZE

PROTECTIVE
COVER
ASSEMBLY

ELECTRICAL
CONNECTIONS

ENCAPSULATED
HEATER
ASSEMBLY

REFERENCE
COLLECTOR

CERAMIC
INSULATORS

FIGURE 1

THERMIONIC DIODE PRESSURE TRANSDUCER

PLASMA-SPRAYED JOINTS

MOLYBDENUM
HEATER POWER
LEADS

ALUMINA COIL FORM

BARE
TUNGSTEN
HEATER

0.105"

0.040"

MOLYBDENUM
HEATER POWER
LEADS

0.125"

FIGURE 2

BASIC HEATER ASSEMBLY

← 0.140" →

CAVITY FOR EMITTER

MOLYBDENUM
EMITTER
HOUSING

ALUMINA
INSULATING
PLATES

BARE TUNGSTEN HEATER

ALUMINA
COIL FORM

ALUMINA INSULATING
PLATES

CAVITY FOR EMITTER

MOLYBDENUM
HEATER POWER
LEADS

FIGURE 3
HEATER ASSEMBLY BEFORE CASTING

29

TUNGSTEN/YTTRIA METALLIZING

MOUNTING POSTS FOR
ENCAPSULATED HEATER
ASSEMBLY

0.145"

REFERENCE COLLECTOR
SURFACE
(TUNGSTEN/YTTRIA)

INSULATORS
FOR HEATER
POWER LEADS

TUNGSTEN/YTTRIA
METALLIZING

LUCALOX CERAMIC

FIGURE 4

CERAMIC BASE

EMITTER

EMITTER HOUSING

ENCAPSULATED HEATER

METAL-CERAMIC JOINT

METAL-CERAMIC JOINT

HEATER LEADS

EMITTER

REFERENCE COLLECTOR SURFACE

METAL-CERAMIC JOINT

METAL-CERAMIC JOINT

Nb-1zr SUPPORT PLATE

REFERENCE COLLECTOR LEAD

FIGURE 5

ENCAPSULATED HEATER-CERAMIC BASE-SUPPORT PLATE ASSEMBLY

31

HEATER MOUNTINGS (ELECTRICAL AND MECHANICAL) (4)

HEATER TERMINALS (4)

VACUUM FLANGE

ELECTRIC HEATERS (2)

HEATER SHIELDS (3)

TEST CAPSULE SUPPORT PLATE

THERMOCOUPLE PORTS (2)

HEATER ELECTRICAL TERMINALS (4)

HEATER TERMINAL COOLING WATER FITTINGS (4)

FIGURE 6

COMPATIBILITY TEST CHAMBER-INTERNAL COMPONENTS

32

CHAMBER
VACUUM
GAUGE

CHAMBER AND TERMINAL
COOLING WATER INLET

VIEWING
WINDOW

OUTLET

POWER
LEADS FOR
HEATERS

TO ION PUMP

CHAMBER COOLING
WATER COILS

FIGURE 7

COMPATIBILITY TEST CHAMBER

FIGURE 8

COMPATIBILITY TEST CAPSULES

CASE I - CANTILEVER, END LOAD

CASE II - END COUPLE, SUPPORTED PIVOT

CASE III - END COUPLE AND END LOAD

FIGURE 9

DIFFERENTIAL PRESSURE TRANSDUCER CONFIGURATIONS

LIST OF REFERENCES

1. Cassano, Anthony J.: Fourth Quarterly Report,
       Pressure Measuring Systems for Closed Cycle
       Liquid Metal Facilities. NASA CR-54371,
       April 5, 1965

2. Engdahl, R. E., ed.: Second Quarterly Report
       Pressure Measuring Systems for Closed Cycle
       Liquid Metal Facilities. NASA CR-54193,
       September 28, 1964.

3. Cassano, Anthony J.: Third Quarterly Report,
       Pressure Measuring Systems for Closed Cycle
       Liquid Metal Facilities. NASA CR-54276,
       December 28, 1964

4. Roark, Raymond J.: Formulas for Stress and
       Strain. Third ed., McGraw-Hill Book.Co.,
       Inc., 1954, p.100.

See also

Engdahl, R. E., ed.: First Quarterly Report
    Pressure Measuring Systems for Closed Cycle
    Liquid Metal Facilities. NASA CR-5414C
    July 2, 1964.

DISTRIBUTION LIST

Recipient	No. of Copies
National Aeronautics & Space Administration Washington, D. C. 20546 Attention: J. J. Lynch, Code RNP	1
U. S. Atomic Energy Commission Washington, D. C. 20545 Attention: SNAP-50/SPUR Program Office Herbert D. Rochen	1
National Aeronautics & Space Administration Ames Research Center Moffett Field, California 94035 Attention: Library	1
National Aeronautics & Space Administration Goodard Space Flight Center Greenbelt, Maryland 20771 Attention: Library	1
National Aeronautics & Space Administration Jet Propulsion Laboratory 4800 Oak Grove Drive Pasadena, California 91103 Attention: Library J. P. Davis	1 1
National Aeronautics & Space Administration Langley Research Center Langley Station Hampton, Virginia 23365 Attention: Library	1

DISTRIBUTION LIST (CONTINUED)

Recipient	No. of Copies
National Aeronautics & Space Administration	
Lewis Research Center	
21000 Brookpark Road	
Cleveland, Ohio 44135	
Attention: Library (3-7)	1
Technology Utilization Office (3-16)	1
Office of Reliability & Quality Assurance (500-203)	1
Dr. B. Lubarsky (500-201)	1
R. F. Mather (500-309)	1
R. N. Weltmann (500-309)	2
R. R. Miller (500-201)	1
H. P. Odom (500-201)	1
M. J. Saari (500-202)	1
C. C. Gettelman (77-1)	1
D. R. Englund, Jr. (77-1)	1
Dr. L. Rosenblum (106-1)	1
H. H. Christenson (5-3)	1
J. E. Dilley (500-309)	1
M. O. Dustin (54-1)	1
E. A. Koutnik (500-201)	1
R. A. Lindberg (500-309)	1
Report Control Office (5-5)	1
National Aeronautics & Space Administration	
George C. Marshall Space Flight Center	
Huntsville, Alabama 35812	
Attention: Library	1
National Aeronautics & Space Administration	
Scientific & Technical Information Facility	
P. O. Box 5700	
Bethesda, Maryland 20014	
Attention: NASA Representative	6

Recipient	No. of Copies
Air Force Systems Command	
Aeronautical Systems Division	
Wright-Patterson AFB, Ohio 45433	
Attention: Library (EWABE)	1
C. H. Armbruster (APIP-1)	1
J. W. Zmurk (APIP-1)	1
G. E. Thompson (APIP-1)	1
G. Sherman (API)	1
U. S. Atomic Energy Commission	
CANEL Project Office	
P. O. Box 1102	
Middletown, Connecticut 06458	
Attention: W. H. Pennington	1
U. S. Atomic Energy Commission	
Germantown, Maryland	
Attention: Col. E. L. Douthett	1
Argonne National Laboratory	
P. O. Box 299	
Lemont, Illinois 60439	
Attention: Library	1
Brookhaven National Laboratory	
Upton, New York 11973	
Attention: Library	1
Dr. O. E. Dwyer	1
Oak Ridge National Laboratory	
Oak Ridge, Tennessee 37831	
Attention: Dr. H. W. Hoffman	
Y-12 Site, Building 9204-1	1
W. C. Thurber	
Metal & Ceramics Division	1

Recipient	No. of Copies
Aerojet-General Corporation SNAP-8 Division Azusa, California 91703 Attention: Dr. Robert Gordon	1
Aerojet General Nucleonics P.O. Box 77 San Ramon, California 94583 Attention: K. P. Johnson	1
AiResearch Manufacturing Company Division of the Garrett Corporation 402 South 36th Street Phoenix, Arizona 85034 Attention: J. Dennan	1
The Bendix Corporation Research Laboratories Division Southfield, Michigan Attention: G. A. Rosselot	1
Data Sensors, Incorporated 13112 Crenshaw Boulevard Gardena, California Attention: J. M. DeStefano	1
Electro-Optical Systems, Inc. 300 N. Halstead Street Pasadena, California 91107 Attention: Dr. J. M. Teem J. C. Sanchez	1 1
Endevco Corporation 801 South Arroyo Parkway Pasadena, California Attention: Library	1

Recipient	No. of Copies
Ford Instrument Company Division of the Sperry Rand Corporation 31-10 Thomson Avenue Long Island City 1, New York Attention: Don Gertz	1
General Electric Company Missile and Space Division Space Power & Propulsion Section Cincinnati, Ohio 45215 Attention: E. Schnetzer	1
General Motors Corporation Allison Division P. O. Box 894 Indianapolis, Indiana 46206 Attention: Dr. F. G. Meyers	1
Gulton Industries, Incorporated Instrumentation Division 212 Durham Avenue Metuchen, New Jersey Attention: Dr. W. Welkowitz	1
Schaevitz-Bytrex Corp. 50 Hunt Street Newton, Massachusetts 02158 Attention: Malcolm Green	1
Materials Research Corporation Orangeburg, New York 10962 Attention: Dr. G. T. Murray	1
McGraw-Edison Company Thomas A. Edison Research Laboratory West Orange, New Jersey Attention: M. D. Bowers	1

Recipient	No. of Copies
MSA Research Corporation Callery, Pennsylvania 16024 Attention: Dr. R. C. Werner	1
North American Aviation, Incorporated Atomics International Division P. O. Box 309 Canoga Park, California 91303 Attention: Dr. L. Bernath Library	1 1
North American Aviation, Incorporated Rocketdyne Division 6633 Canoga Avenue Canoga Park, California 91303 Attention: V. R. Degner	1
Physical Sciences Corporation 314 East Live Oak Avenue Arcadia, California Attention: J. Howard Pitt	1
Trans-Sonics, Incorporated Pressure Instrumentation Division Burlington, Massachusetts Attention: R. E. Jackson	1
United Aircraft Corporation Pratt & Whitney Aircraft Division East Hartford, Connecticut 06108 Attention: Library	2
West Coast Research Corporation 2102 South Sepulveda Boulevard Los Angeles, California 90025 Attention: H. M. Spivack	1

Recipient	No. of Copies
Westinghouse Electric Corporation Astronuclear Laboratory P. O. Box 10864 Pittsburgh, Pennsylvania 15236 Attention: W. D. Pouchot	1
Westinghouse Electric Corporation Aerospace Electric Division Lima, Ohio 45801 Attention: P. E. Kueser	1